Comparison of the Measuring Devices during the Evaluating of Circularity Deviation

Robert Čep^{1,*}, Šárka Malotová¹, Vladimír Vrba¹

¹ Department of Machining, Assembly and Engineering Metrology, Faculty of Mechanical Engineering, VŠB-TU Ostrava, 17. listopadu 15/2172, Ostrava-Poruba, 708 33, Czech Republic

Abstract: The contribution deals with the comparison of devices for measuring and evaluation of circularity of deviation. The circularity of deviation was measure on the specific component using in practice. The deviation was measured with using standard circular gauge Formtester MMQ 44 CNC and 3D measuring device SCAN Max. The results were statistical evaluated in tables. Part of the results is the technical and economic evaluation too.

Keywords: *metrology, circularity deviation, 3D device, measurement.*

1. Introduction

The main aim of machining process is production removable components with maximum functionality and adequately costs. Each part needs accuracy production with predefined dimensions, geometry and surface adjustment. With development more accuracy machine tools are allowed to make very small components in tolerances. Geometrical specification of component in engineering and metrology are essential for proper assembly and safety of manufactured parts. All of them have cylindrical and circular shape and control is necessary for safety state of all system. The tolerances of shape (e.g. circularity) are checked with using 3D measuring machine and devices, which allow repeatability, accuracy and reliability of measuring. [1, 2]

The circularity deviation is defined by standard ČSN EN ISO 1101 – Geometrical specification of components – geometrical tolerance – The tolerance of shape, orientation, location and throwing. Real surfaces show deviation given nominal values. In this tolerance field must be the actual element. The circularity deviation is the biggest perpendicular distance, which is measured to envelope circle (see fig. 1). Circularity deviations are measured with special touch device, which rotates around fixed component, respectively conversely or used CMM. The circularity deviation (Δ) of points of actual profile of enveloping circle is the biggest distance. These deviations together with surface roughness significantly influence the level of vibration of rotating parts, hence they need to be properly diagnosed and improving the whole process of production. (Figure 1) [3]

2. Check selected component

The checked component is part of system using in automobile SCANIA P420/R420 into motor part – pulley (figure 2). The component was machined and then quenching into a salt bath. Then component was moved to laboratory of metrology, where was measured tolerances include circularity deviation. The nominal value of ring was $\varnothing$ 15 mm. The control of circularity was measured with

* Corresponding author: Robert Čep, E-mail address: robert.cep@vsb.cz

Fig. 1: *Circularity deviation* [1].

Fig. 2: Measured component [4].

using circular gauge Formtester MMQ 44 CNC and 3D measurement device SCAN Max. [4]

Deviation was measured with using standard circular gauge Formtester MMQ 44 CNC on principle Least squares method (LSC). Before the measuring was created working operation. The component was fixed into chuck and measured by the ball arm. It is very accuracy device for checking of deviation of shape and location according to standard ČSN EN ISO 1101. The device is composed from sliding arm,

measuring range (X – Z: 180 – 500) mm, rotary table, measuring touch rotatable around the 360° axis and PC with software (figure 3). This device offers double the accuracy in roundness measurements, improved straightness measurements, and faster measuring speeds. The system will measure parts up to 620 mm in diameter and 640 mm in height. [5]

Fig. 3: *Formtester MMQ 44 CNC* [5].

Fig. 4: *SCAN Max – 3D device [6].*

Then the circularity deviation was measured with using SCAN Max 3D device (figure 4). SCAN Max is Carl Zeiss Coordinate Measurement Machine (CMM). Partially manual control along with active scanning means high precision combined with ease of use. SCAN Max is extremely small in relation to its wide range of applications and can be easily placed out of the workshop. It has Scan Ware operating system with ball probes with diameters 2, 3, 5, 8mm, plate diameter 15mm and thickness 1mm, 3mm diameter

roller probe. Measuring range is (850 x 450 x 450) mm. The both measurement processed under metrological conditions in accredited laboratories.

3. Measured values

The measuring of circularity of deviation (LSC) proceeded under constant conditions – ambient temperature 20,1°C and number of measuring was thirty. The results were took into table 1 and then was evaluated uncertainty type A and B.

For calculation of standard uncertainty of type B, was took into the expanded uncertainties of

Tab. 1: Measured values.

Circularity deviation – Coordinate Measuring Device				Circularity deviation – Circular gauge			
No.	Δ [mm]	No.	Δ [mm]	No.	Δ [mm]	No.	Δ [mm]
1	0,0034	16	0,0027	1	0,00076	16	0,00081
2	0,029	17	0,0026	2	0,00078	17	0,00079
3	0,0027	18	0,0023	3	0,00078	18	0,0008
4	0,0029	19	0,0027	4	0,00079	19	0,00081
5	0,0031	20	0,0032	5	0,00077	20	0,00078
6	0,0028	21	0,0027	6	0,00078	21	0,00079
7	0,003	22	0,0034	7	0,0008	22	0,00079
8	0,0027	23	0,0023	8	0,00081	23	0,00077
9	0,0023	24	0,0027	9	0,00079	24	0,00077
10	0,003	25	0,0024	10	0,00077	25	0,00078
11	0,0029	26	0,0028	11	0,00076	26	0,00077
12	0,0027	27	0,0032	12	0,00082	27	0,00079
13	0,0028	28	0,002	13	0,00084	28	0,00078
14	0,003	29	0,0029	14	0,00082	29	0,00079
15	0,0026	30	0,0023	15	0,0008	30	0,00079
$\bar{\Delta}$		0,003637 mm		$\bar{\Delta}$		0,000789 mm	
$s(\bar{\Delta}) = u_A(\bar{\Delta})$		0,000862 mm		$s(\bar{\Delta}) = u_A(\bar{\Delta})$		0,00001843 mm	

the 3D device and temperature correction δT . The temperature deviation of measuring component from the ambient temperature stayed $\pm 0,2$ K. The ambient temperature in laboratory was $\pm 0,5$ K. We

did not take into the influence of uncertainty of ring. Temperature influences of laboratory δT_A and ring δT_X was ignored due to very low values. The results of measurement are shown in table 2.

Tab. 2: Results of measurement.

Quantity		Circularity deviation – Coordinate Measuring Device	Circularity deviation – Circular gauge
Laboratory	δT_A	insignificant	insignificant
Bearing ring	δT_X	insignificant	insignificant
Temperature correction	δT	insignificant	insignificant
Uncertainty of device	δT_i	1,7 μm	0,16 μm
Total uncertainty	U_B	1,7 μm	0,16 μm
Combined uncertainty	U_c	1,7 μm	0,16 μm
Extended uncertainty	U	3,4 μm	0,32 μm
Results of measurement	Δ	(0,00364 $\pm$ 0,00340) mm	(0,00079 $\pm$ 0,00032) mm
Tolerance $\pm$ 0,005 mm		UNSUITABLE	SUITABLE

The steel ring was measured thirty times by the both devices and statistically evaluated. The circular gauge was suitable and satisfied the allowed tolerance.

3. Technical and economic evaluation - conclusion

From the calculations it can be stated that the 3D device SCAN Max is only measuring instrument and in comparison with laboratory circular gauge Formtester MMQ 44 CNC, which is used for measuring of deviations of shape a location and its accuracy is higher than 3D device SCAN Max. But in our condition where is tolerance $\pm$ 50 μm we could use 3D device for control the circularity of deviation, because it is only inter-operational control. Determination of time saving was calculated on the principle of number seven orders of 2 pieces per one work shift. In table 3, there is show time savings for devices.

The aim of the work was to compare measuring device during evaluation of circularity deviation of

selected components from practice. The measuring devices were circular gauge Formtester MMQ 44 CNC and 3D measuring device SCAN Max, which are used in workroom like a measuring instrument. For the conditions in workroom and inter-operational control, they could use less accuracy device, because the component will be further machined and on the economic page it is better to use 3D device.

Tab. 3: Time saving.

Device	Number of samples per 1 work shift [ks]	Number of standard minute per 1 sample [Nmin]	Number of standard min per 1 day [Nmin/day]	Time savings for 1 year [day] – 3D device
3D device	14	6	84	9,52
Circular gauge	14	10	140	
Time saving with using 3D device for measuring and evaluation of circularity deviation was 9.5 days.				

Tab. 4: Financial saving.

Device	Number of samples per 1 work shift [ks]	Hour rate [Kč]	Kč per work shift [Kč]	Financial savings for 1 day [Kč] – 3D device	Financial savings for 1 year [Kč] – 3D device
3D device	14	120	168	205,33	50 305,85
Circular gauge	14	160	373,33		
In table 4 is counted off financial saving with using 3D device per one year, which is 50 306 CZK.					

References and Notes

- [1] ČSN ISO 230 - 4. Test code for machine tools – Part 4: Circular tests for numerically controlled mach. Czech standard. 2010.
- [2] Technical drawings - Geometrical Tolerancing, ISO 1101, (1983)12-01
- [3] SOUZA, C.C. at al. A CONTRIBUTION TO THE MEASUREMENT OF CIRCULARITY AND CYLINDRICITY DEVIATIONS. Brazilian Congress of Mechanical Engineering. Brazil, 2011, (21), 791-800.
- [4] IRIŠEK, D. Roundness Measuring at Coordinating Measuring Machines. Ostrava: Department of Working and Assembly, Faculty of Mechanical Engineering, VŠB-Technical University of Ostrava, 2009, 42 s. Thesis head: Vrba Vladimír.
- [5] Mahr Federal Introduces New And Improved Formtester. MetrologyWorld.com [online]. 2005 [cit. 2017-07-26]. Available from: <https://www.metrologyworld.com/doc/mahr-federal-introduces-new-and-improved-form-0001>
- [6] Zeiss, Carl. Scanmax and probe. Available from: <http://www.ufix.se/scanmax.html>.

Biographical notes

Assoc. Prof. Robert Čep, Ph.D., MSc. (born 1977) got the title at VSB-Technical University of Ostrava in field Mechanical Engineering. He specializes about machining and engineering metrology. Currently, he is vice-dean for PhD studium, external workplaces and personal development. He deals with machining and its optimalization, testing of cutting tools, surface integrity and measuring accuracy of machine tools. Šárka Malotová, MSc. (born 1990) is Ph.D. student at Dpt. Machining, Assembly and Engineering Metrology. She deals with measuring of surface topography and testing of cutting tool. Her dissertation thesis is about the evaluation of durability of cutting tools during parting-off of stainless steels. Assoc. Prof. Vladimír Vrba, CSc, MSc. is emeritus professor at pt. Machining, Assembly and Engineering Metrology. His field is Engineering Metrology.